

GOTECH "the best partner" to raise your competition

GT-6011-2-1 芯紙豎壓試驗夾具

Fluted Crush Specimen Holder (Cfc)

TAPPI-T824

適用於瓦楞芯紙的豎壓強度測試。採用此治具製作瓦楞芯紙豎壓強度測試之試樣。

This apparatus is used as a specimen holding device in corrugated fluted crush test to evaluate the ability of corrugating medium. A fluted specimen of corrugating medium is held by the specimen holder and compressed by a compression tester.

齒高 Height of teeth	$4.75 \pm 0.05 \text{ mm}$
齒距 Tooth spacing	$8.50 \pm 0.05 \text{ mm}$
齒寬(厚度) Width of teeth	$6.35 \pm 0.05 \text{ mm}$
體積 Dimensions(W×D×H)	$190 \times 75 \times 13 \text{ mm}$



瞭解更多產品資訊，請登錄我司網站 www.gotech.biz